


(COMPRESSION) DISINTEGRATION TEST REPORT

Product Name	AMOXYCILLIN TRIHYDRATE D.P.		
Batch No.	BN-DT2104	Department	Compression
BFG Code No.	BAC-993100	Date	21.04.2026
No. of Samples	6	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	EQ8800	Disintegration Tester Code	DT-80006
Media	Water	Temperature Range (°C)	35.0 – 39.0
Cubicle Name	Compression_1		

Test Results

Min. Temperature Jar(°C)	36.5	Max. Temperature Jar(°C)	37.0
Min. Disintegration Time	00:00:23	Max. Disintegration Time	00:02:37
Set Temperature	37		



Sample No.	Start Time	End Time	Halt Duration	Disintegration Time
1	16:50:24	16:50:47	00:00:00	00:00:23
2	16:50:24	16:51:23	00:00:00	00:00:59
3	16:50:24	16:51:52	00:00:00	00:01:28
4	16:50:24	16:51:07	00:00:00	00:00:43
5	16:50:24	16:50:58	00:00:00	00:00:34
6	16:50:24	16:53:01	00:00:00	00:02:37

Result: Complies

Performed by : Durgesh Shah

Date : 21.04.2026