

AUDIT TRAIL FOR BALANCE



DATE	TIME	USER ID/USER NAME	ACTION	BALANCE CODE NO.	OLD PARAMETER	NEW PARAMETER	OLD WEIGHT	NEW WEIGHT	REMARK
01/09/26	16:10	User Id: 4005 User Name: Tushar	Activate	HC-105	NA	NA	NA	NA	OQ-SRS-5
01/09/26	15:36	User Id: 4006 User Name: Nikhil	Verify	HC-105	NA	NA	NA	NA	OQ-SRS-5
01/09/26	14:46	User Id: 4001 User Name: Rahul	Add	HC-105	NA	Serial No.: B445231906 , Balance Code No.: HC-105 , Make: Mettler , Model: XS205DU , Select Unit: g , Least Count: 0.0001 , Minimum Capacity: 0.0001 , Maximum Capacity: 220.0000 , Minimum Operating Range: 0.0100 , Maximum Operating Range: 209.0000 , Calibration Start date: 2026-09-01 , Calibration Duration in Months: 12 , Set Reminder: 30	NA	Daily(P): Standard Weight:0.01,(-)Tolerance:0.0099, (+)Tolerance:0.0101 Standard Weight:209, (-)Tolerance:206.91,(+)Tolerance:211.09 Monthly(P): Standard Weight:0.01,(-)Tolerance:0.0099, (+)Tolerance:0.0101 Standard Weight:209, (-)Tolerance:206.91,(+)Tolerance:211.09 Standard Weight:44,(-)Tolerance:43.56,(+)Tolerance:44.44 Standard Weight:88,(-)Tolerance:87.12, (+)Tolerance:88.88 Standard Weight:132, (-)Tolerance:130.68,(+)Tolerance:133.32 Standard Weight:176,(-)Tolerance:174.24,(+)Tolerance:177.76 Linearity(P): Standard Weight:0.01,(-)Tolerance:0.0099, (+)Tolerance:0.0101 Standard Weight:209, (-)Tolerance:206.91,(+)Tolerance:211.09 Standard Weight:176,(-)Tolerance:174.24,(+)Tolerance:177.76 Standard Weight:66,(-)Tolerance:65.34, (+)Tolerance:66.66 Standard Weight:110, (-)Tolerance:108.9,(+)Tolerance:111.1 Eccentricity(P): Standard Weight:110,(-)Tolerance:108.9, (+)Tolerance:111.1 Repeatability(P): Standard Weight:11,(-)Tolerance:10.89,(+)Tolerance:11.11	OQ_SRS_5