

Balance ID no:	CE-159	Location:	Production	Operating Range :	0.0100 - 304.0000 gm
Balance Capacity :	0.0001 - 320.0000 gm	Least Count :	0.0001 gm	Calibration Frequency	12 day
Weight Box / Standard Weight ID No.	TAB-29			Valid Upto:	21/11/26
Date of Calibration :	27/03/26			Next Calibration Due on:	27/03/26
Reason for Calibration	☐ Schedule ☐ Non Schedule _____				

1. Surface level check

Surface level check	Observation
Check the surface level on the level (Bubble) indicator and ensure that the bubble should be in Center	Satisfactory
Calibration Done By : testing.8(Kiran)	

2. Sensitivity and Linearity

Weight Selection for calibration	Selected weights	Acceptance Criteria	Actual Observation (Center position)	
Minimum of Entire operation range	0.0100 gm	Actual weight = ± 0.5 % of Selected standard Weight (For Dispensing activity)	0.0100 gm	
30% of Balance capacity	96.0000 gm		Actual weight = ± 0.1 % of actual mass * of Selected standard weight (For Dispensing activity for balance having capacity 220 g or below with least count 0.00001 g or better)	96.0000 gm
50% of Balance capacity	160.0000 gm			160.0000 gm
80% of Balance capacity	256.0000 gm			256.0000 gm
Maximum of Entire operating range	304.0000 gm	Actual weight = ± 1.0 % of Selected standard weight (For other than Dispensing activity)	304.0000 gm	
Calibration done By :		testing.8(Kiran)		
* Actual mass of standard weights:				

3. Eccentricity test

Weight Selection for calibration	Selected weights	Acceptance Criteria	Actual Observation	
50% of Balance capacity	160 gm	Actual weight = ± 0.5 % of Selected standard Weight (For Dispensing activity) Actual weight = ± 0.1 % of actual mass * of Selected standard weight (For Dispensing activity for balance having capacity 220 g or below with least count 0.00001 g or better) Actual weight = ± 1.0 % of Selected standard weight (For other than Dispensing activity)	1 Position	160.0000 gm
			2 Position	150.0000 gm
			3 Position	
			4 Position	
			Center	
Calibration done By :			testing.8(Kiran)	
* Actual mass of standard weights :				

Remarks: _____

Conclusion : The Calibration of instrument found satisfactory / Not satisfactory

Done By/ Date: testing.8(Kiran) / 27/03/26

Verified By / Date: