



In Process Quality Control Test Report

Friability Report (Regular)

Product Name	Compression OQ		
Stage	NA	Batch No.	OQTESTIPQC
Test Start Date/Time	16/07/2026 14:21	Test End Date/Time	16/07/2026 14:25
Instrument ID No.	T-526	Cubicle No.	4
Comments			

Limits :	
Friability (%) NMT 1 % W/W	
Set RPM :	25
Set Count :	100

Calculation	
Drum 1 (NA)	
Weight before test (W1)	100.0000
Weight after test (W2)	99.4500
Formula : $\frac{[W1 - W2] \times 100}{W1}$	
Breakage	NA
Friability %	0.55

Performed by	4001 (Rahul Lad)
--------------	------------------