

Balance Verification Record (Daily/Monthly)

Unit	gm		
Balance Code No.	T-555	Department	NA
Month / Year	Mar / 2026	Capacity (gm)	220.000
Make	Mettler	Least Count (gm)	0.001
Model	ML203T/A00		
Std. Weight Box ID	T-515		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (Std Wt +/- Tolerance) (gm)
Lower Weight	0.020	± 0.001	0.019 - 0.021
Middle 1 Weight	1.000	± 0.001	0.999 - 1.001
Upper Weight	100.000	± 0.100	99.900 - 100.100

Date	Time	Spirit Level(Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)			Performed By	Checked By	Remark
				0.020	1.000	100.000			
03.04.2026	17:18:24	OK	Nil	0.020	1.000	100.000	testing.8	testing.7	Ok