

Balance Verification Record (Daily/Monthly)

Unit	gm		
Balance Code No.	T-554	Department	QC
Month / Year	Jan - 26	Capacity (gm)	220.000
Make	Mettler	Least Count (gm)	0.001
Model	ML203T/A00		
Std. Weight Box ID	T-515		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (Std Wt +/- Tolerance) (gm)
Lower Weight	0.020		0.019 - 0.021
Middle 1 Weight	1.000		0.999 - 1.001
Upper Weight	100.000		99.900 - 100.100

Date	Time	Spirit Level(Ok/Not Ok)	Zero Error	Standard Weights			Performed By	Checked By	Remark
				0.020	1.000	100.000			
01.04.2026	15:01:35	0	0	0.020	1.000	100.000	testing.8		Ok