

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	gm		
Balance Code No.	T/249	Department	Compression
Month / Year	May / 2026	Capacity (gm)	210.000
Operating Range (gm)	0.019 - 168.168	Least Count (gm)	0.001
Std. Weight Box ID	WB/GB/WB-002,WTB002		
Certificate Number / Validity	SCL/2022-893,100 / 07.01.2027, 14.12.2026		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (gm)
Lower Weight	0.020	± 0.001	0.019 - 0.021
Middle 1 Weight	4.200	± 0.004	4.196 - 4.204
Upper Weight	42.000	± 0.042	41.958 - 42.042

Date	Time	Cleanliness (Clean / Unclean)	Spirit Level (Ok/Not OK)	Zero Error	Standard Weights for Verification (gm)			Calibration Status	Done By	Checked By	Remark
					0.020	4.200	42.000				
01.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
02.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
03.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
04.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
05.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
06.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
07.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
08.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
09.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
10.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD

Date	Time	Cleanliness (Clean / Unclean)	Spirit Level (Ok/Not OK)	Zero Error	Standard Weights for Verification (gm)			Calibration Status	Done By	Checked By	Remark
					0.020	4.200	42.000				
11.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
12.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
13.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
14.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
15.05.2026	11:47:23	Clean	OK	No Zero Error	0.020	4.200	42.000	Within Limit	22	22	Ok
16.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
17.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
18.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
19.05.2026	10:11:54	Clean	OK	No Zero Error	0.020	4.199	41.996	Within Limit	Tantra	22	Ok
20.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
21.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
22.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
23.05.2026	13:48:34	Clean	OK	No Zero Error	0.020	4.200	42.000	Within Limit	Tantra	22	Ok