

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	gm		
Balance Code No.	PC/GB/EBL-002	Department	IT
Month / Year	May / 2026	Capacity (gm)	420.000
Operating Range (gm)	0.020 - 340.000	Least Count (gm)	0.001
Std. Weight Box ID	WB/GB/WB-002		
Certificate Number / Validity	SCL/2022-893 / 07.01.2027		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (gm)
Lower Weight	0.020	± 0.001	0.019 - 0.021
Upper Weight	340.00	± 0.340	339.660 - 340.340

Date	Time	Cleanliness (Clean / Unclean)	Spirit Level (Ok/Not OK)	Zero Error	Standard Weights for Verification (gm)		Calibration Status	Done By	Checked By	Remark
					Lower Weight (gm)	Upper Weight (gm)				
01.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
02.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
03.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
04.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
05.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
06.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
07.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
08.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
09.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
10.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
11.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA

DAILY VERIFICATION AND ZERO ERROR RECORD

Date	Time	Cleanliness (Clean / Unclean)	Spirit Level (Ok/Not OK)	Zero Error	Standard Weights for Verification (gm)		Calibration Status	Done By	Checked By	Remark
					Lower Weight (gm)	Upper Weight (gm)				
12.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
13.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
14.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
15.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
16.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
17.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
18.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
19.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
20.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
21.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
22.05.2026	NA	NA	NA	NA	NA	NA	NA	NA	NA	NA
23.05.2026	11:25:04	Clean	OK	No Zero Error	0.020	-	Out Of Limit	Tantra	22	No Activiy
23.05.2026	11:27:13	Clean	OK	No Zero Error	0.020	-	Out Of Limit	Tantra		No Activiy