



CALIBRATION OF VERNIER

INSTRUMENT CODE : TESTVER

CALIBRATED ON : 26/05/2026

NEXT CALIBRATION DUE ON : 25/06/2026

Dimensions of Standard Blocks :

Check the dimensions of the standard blocks of 15, and 25 mm and record the observations in following table.

Sr. No	Parameter	Standard	Acceptance Criteria	Observation	Checked by
1	2.00	2.00	1.99 2.01	2.00	
2	5.00	5.00	4.99 5.01	5.00	
3	10.00	10.00	9.99 10.01	10.00	

Remark : Ok

CONCLUSION : The instrument is calibrated as per written procedure and complies with acceptance criteria.

Calibration done by : 3

Check By : TEST1

Date :

Date : 26/05/2026