

AUDIT TRAIL FOR BALANCE

DATE	TIME	USER ID	ACTION	BALANCE CODE NO.	OLD VALUES	NEW VALUES	OLD WEIGHT	NEW WEIGHT	REMARK
07.01.2026	11:54:01	User ID:123,User Name:yogita	Activate	BAL-070126	NA	NA	NA	NA	E

AUDIT TRAIL FOR BALANCE

DATE	TIME	USER ID	ACTION	BALANCE CODE NO.	OLD VALUES	NEW VALUES	OLD WEIGHT	NEW WEIGHT	REMARK
07.01.2026	11:45:59	User ID:22,User Name:Tantra	Add	BAL-070126	NA	Balance Code No.: BAL-070126,Make: METTLER TOLEDO,Model: MS303S,Serial No.: NA,Unit: g,Least Count: 0.001,Maximum Capacity: 320.000,Minimum Capacity: 0.001,Maximum Operating Range: 250.000,Minimum Operating Range: 0.020,Periodic Data: Calibration Date:07-01-2026,Duration :5,Reminder:4,Balance Type:Analytical	NA	Daily: Standard Weight:0.020,(-)Tolerance:0.010,(+)Tolerance:0.030 Standard Weight:0.500,(-)Tolerance:0.489,(+)Tolerance:0.510 Periodic: Standard Weight:0.020,(-)Tolerance:0.010,(+)Tolerance:0.030 Standard Weight:0.500,(-)Tolerance:0.489,(+)Tolerance:0.510 Standard Weight:64.000,(-)Tolerance:63.999,(+)Tolerance:64.100 Standard Weight:120.000,(-)Tolerance:119.000,(+)Tolerance:120.100	2