

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	X		
Balance Code No.	balance-new-1012	Department	QC
Month / Year	December / 2025	Capacity (g)	320.000
Make	METTLER TOLEDO	Least Count (g)	0.001
Model	ML303		
Std. Weight Box ID	NA		
Std. Weight Box Certificate No.	NA		
Std. Weight Box Valid Upto Date	NA		

Standard Weights for Verification (g)		Tolerance (g)	Accuracy Limits (Std Wt +/- Tolerance) (g)
Lower (20.e) / (5% of a) / Minimum Capacity of the checkweigher / weighing balance	0.020	(± 0.010)	0.010 - 0.030
Upper (80% of a)	0.500	(± 0.050)	0.450 - 0.550

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (g)		Performed By	Checked By	Remark
				0.020	0.500			
01.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
02.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
03.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
04.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (g)		Performed By	Checked By	Remark
				0.020	0.500			
05.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
06.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
07.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
08.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
09.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
10.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
11.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
12.12.2025	10:05	OK	Nil	0.020	0.500	yogita	Tantra	OK
13.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
14.12.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
15.12.2025	10:22	OK	Nil	0.020	0.500	yogita	Tantra	OK