

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	X		
Balance Code No.	Balance-1111	Department	NA
Month / Year	November / 2024	Capacity (gm)	350.000
Make	METTLER TOLEDO	Least Count (gm)	0.001
Model	ML303		
Std. Weight Box ID	WBI_1		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (Std Wt +/- Tolerance) (gm)
Lower Weight	0.020	(± 0.002)	0.018 - 0.022
Upper Weight	10.000	($- 0.104 + 0.123$)	9.896 - 10.123

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)		Performed By	Checked By	Remark
				0.020	10.000			
01.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
02.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
03.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
04.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
05.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD



Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)		Performed By	Checked By	Remark
				0.020	10.000			
06.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
07.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
08.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
09.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
10.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
11.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
12.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
13.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
14.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
15.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
16.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD



Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)		Performed By	Checked By	Remark
				0.020	10.000			
17.11.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
18.11.2024	15:16	OK	Nil	0.020	10.020	Dipti	yogita	OK