

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	X		
Balance Code No.	BAL1707	Department	QC
Month / Year	August / 2025	Capacity (g)	320.000
Make	METTLER TOLEDO	Least Count (g)	0.001
Model	ML303		
Std. Weight Box ID	WB16072025		

Standard Weights for Verification (g)		Tolerance (g)	Accuracy Limits (Std Wt +/- Tolerance) (g)
Lower (20.e) / (5% of a) / Minimum Capacity of the checkweigher / weighing balance	10.000	(± 1.000)	9.000 - 11.000
Upper (80% of a)	30.000	(± 1.000)	29.000 - 31.000

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (g)		Performed By	Checked By	Remark
				10.000	30.000			
16.08.2025	17:06	OK	Nil	10.000	30.000	kumkum	Tantra	OK
17.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
18.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
19.08.2025	10:29	OK	Nil	10.000	30.000	kumkum	Tantra	OK

DAILY VERIFICATION AND ZERO ERROR RECORD

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (g)		Performed By	Checked By	Remark
				10.000	30.000			
20.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
21.08.2025	11:51	OK	Nil	10.000	30.000	kumkum	Tantra	OK
22.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
23.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
24.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
25.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
26.08.2025	15:11	OK	Nil	10.000	30.000	kumkum	Tantra	OK
27.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
28.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
29.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity
30.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (g)		Performed By	Checked By	Remark
				10.000	30.000			
31.08.2025	NA	NA	NA	NA	NA	NA	NA	No Activity