

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	X		
Balance Code No.	Bal_22	Department	QC
Month / Year	October / 2024	Capacity (gm)	320.000
Make	METTLER TOLEDO	Least Count (gm)	0.001
Model	ML203		
Std. Weight Box ID	WBI_1		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (Std Wt +/- Tolerance) (gm)
Lower Weight	0.020	(± 0.001)	0.019 - 0.021
Upper Weight	256.000	($- 0.005 + 0.004$)	255.995 - 256.004

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)		Performed By	Checked By	Remark
				0.020	256.000			
01.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
02.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
03.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
04.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
05.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD



Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)		Performed By	Checked By	Remark
				0.020	256.000			
06.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
07.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
08.10.2024	11:00	OK	Nil	0.020	256.000	Abhishek	Sail	OK
09.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
10.10.2024	09:23	OK	Nil	0.020	256.000	Abhishek	Sail	OK
11.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
12.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
13.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
14.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
15.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
16.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD



Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)		Performed By	Checked By	Remark
				0.020	256.000			
17.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
18.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
19.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
20.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
21.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
22.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
23.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
24.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
25.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
26.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
27.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD



Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)		Performed By	Checked By	Remark
				0.020	256.000			
28.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
29.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
30.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity
31.10.2024	NA	NA	NA	NA	NA	NA	NA	No Activity