

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	X		
Balance Code No.	Balance-121212	Department	QC
Month / Year	October / 2024	Capacity (gm)	350.000
Make	METTLER TOLEDO	Least Count (gm)	0.001
Model	ML303		
Std. Weight Box ID	WBI_1		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (Std Wt +/- Tolerance) (gm)
Lower Weight	0.020	(± 0.002)	0.018 - 0.022
Middle 1 Weight	10.000	(- 0.104 + 0.452)	9.896 - 10.452
Middle 2 Weight	20.000	(± 2.000)	18.000 - 22.000
Upper Weight	30.000	(- 2.000 + 1.000)	28.000 - 31.000

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)				Performed By	Checked By	Remark
				0.020	10.000	20.000	30.000			
01.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
02.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
03.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
04.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD



Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)				Performed By	Checked By	Remark
				0.020	10.000	20.000	30.000			
05.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
06.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
07.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
08.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
09.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
10.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
11.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
12.10.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
13.10.2024	10:26	OK	Nil	0.021	10.000	20.000	30.000	Priti	Mangesh	OK