

DAILY VERIFICATION AND ZERO ERROR RECORD

Unit	X		
Balance Code No.	BAL-0348	Department	NA
Month / Year	April / 2024	Capacity (gm)	320.000
Make	METTLER TOLEDO	Least Count (gm)	0.001
Model	ML303		
Std. Weight Box ID	WT-0953		

Standard Weights for Verification (gm)		Tolerance (gm)	Accuracy Limits (Std Wt +/- Tolerance) (gm)
Lower Weight	0.020	(- 0.002 + 0.003)	0.018 - 0.023
Middle 1 Weight	64.000	(± 2.000)	62.000 - 66.000
Middle 2 Weight	120.000	(± 2.000)	118.000 - 122.000
Upper Weight	150.000	(± 2.000)	148.000 - 152.000

Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)				Performed By	Checked By	Remark
				0.020	64.000	120.000	150.000			
23.04.2024	10:50	OK	Nil	0.020	64.000	120.000	150.000	Manish	Tantra	OK
24.04.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
25.04.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
26.04.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity

DAILY VERIFICATION AND ZERO ERROR RECORD



Date	Time	Spirit Level (Ok/Not Ok)	Zero Error	Standard Weights for Verification (gm)				Performed By	Checked By	Remark
				0.020	64.000	120.000	150.000			
27.04.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
28.04.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
29.04.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity
30.04.2024	NA	NA	NA	NA	NA	NA	NA	NA	NA	No Activity