

FULL SCALE CALIBRATION AND ZERO ERROR RECORD
(Analytical and Precision Weighing Balance)

Unit : X

Department : QC

Month and year : March 2025

Balance Code No. : BAL-0412

Capacity (gm) : 320.000

Make : METTLER TOLEDO

Model : ML203

Std. Weight Box ID. : WT-0415

Least Count (gm) : 0.001

Tolerance: +/- Least count of the balance OR +/- 0.1 % of Standard weight whichever is higher and value should be rounded off on lower side as per display capacity of balance.

Standard weights for calibration (gm)		Tolerance (gm)	Accuracy Limits(Std wt +/- Tolerance) (gm)
Lower	0.020	± 0.002	0.018 - 0.022
Middle(1)	64.000	± 2.000	62.000 - 66.000
Middle(2)	120.000	± 2.000	118.000 - 122.000
Upper	150.000	± 2.000	148.000 - 152.000

Note: 1. Adjust zero error before commencing the weighing operation.

2. Check against standard weights on quarterly.

Date	Time	Spirit Level (Ok/ Not Ok)	Zero Error	Standard weights kept on the balance (gm)		Readings shown (gm)	Remark	Performed by	Checked By
11.03.2025	16:18	OK	Nil	Lower	0.020	0.020	Ok	Manish(C)	Tantra(22)
				Middle(1)	64.000	64.001	Ok	Manish(C)	Tantra(22)
				Middle(2)	120.000	120.001	Ok	Manish(C)	Tantra(22)
				Upper	150.000	150.001	Ok	Manish(C)	Tantra(22)

FULL SCALE CALIBRATION AND ZERO ERROR RECORD
(Analytical and Precision Weighing Balance)

2. Uncertainty test:

Sr. No.	Standard Weight (gm)	Reading displayed on the balance (gm)	Remarks
1	0.020	0.022	Ok
2	0.020	0.022	Ok
3	0.020	0.022	Ok
4	0.020	0.022	Ok
5	0.020	0.022	Ok
6	0.020	0.022	Ok
7	0.020	0.022	Ok
8	0.020	0.022	Ok
9	0.020	0.022	Ok
10	0.020	0.022	Ok
Mean		0.022	-
Standard Deviation (S.D.)		0.0000	-

Uncertainty =
$$\frac{3 \times \text{S.D.}}{\text{Standard Weight (gm)}} = 0.150$$

Acceptance criteria : Value of uncertainty should be NMT 0.001 gm.

Remark : Not Complies

FULL SCALE CALIBRATION AND ZERO ERROR RECORD
(Analytical and Precision Weighing Balance)

3. Repeatability test:

Sr. No.	Standard Weight (gm)	Reading displayed on the balance (gm)	Remarks
1	0.020	0.019	Ok
2	0.020	0.019	Ok
3	0.020	0.020	Ok
4	0.020	0.022	Ok
5	0.020	0.022	Ok
6	0.020	0.020	Ok
7	0.020	0.020	Ok
8	0.020	0.020	Ok
9	0.020	0.020	Ok
10	0.020	0.020	Ok
Mean		0.020	-
Standard Deviation (S.D.)		0.0010	-

$$\text{Repeatability (\%)} = \frac{2 \times (0.41 \times d)}{\text{Standard weight (gm)}} \times 100 = 4.10$$



Acceptance criteria : Repeatability is NMT 0.10 %

Remark : Not Complies

FULL SCALE CALIBRATION AND ZERO ERROR RECORD
(Analytical and Precision Weighing Balance)

4. Eccentricity test :

Note: Adjust zero error before commencing the weighing operation .



Sr No.	Standard Weight (gm)	Position	Reading displayed by the balance (gm)	Deviation from center =X-C	Remark
1	0.020	C	0.022	-----	Ok
2	0.020	C1	0.022	0.000	Ok
3	0.020	C2	0.022	0.000	Ok
4	0.020	C3	0.022	0.000	Ok
5	0.020	C4	0.022	0.000	Ok

Where, X=C1, C2, C3, C4 i.e. different corners of the balance.

Acceptance criteria : Deviation is not more than 0.05%.

Remark: Complies

Comment: _____

Calibration Done By: Manish (C)

Checked By:
(Quality Assurance)

Date : 11.03.2025

Date :