


REGULAR DISINTEGRATION TEST REPORT

Product Name	KROCIN TABLET1		
Batch No.	BT-121212	Department	QC
BFG Code No.	KROCIN TABLET1	Date	18.11.2024
No. of Samples	10	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	None	Disintegration Tester Code No.	ED 3PO
Media	NA	Side	NA
Temperature Range (°C)	34.8 - 38.9		

Test Results

Min. Temperature (°C)	28.6	Max. Temperature (°C)	28.6
Max. Disintegration Time	00:01:51	Set Temperature	38.0
Mode	NA	Jar Detail	A
Basket Type	10 Mesh		



—●— Actual Value —◆— Standard

Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	11:54:10	11:56:01	NA	00:01:05	00:01:51

Result: **Not Complies**
test

Performed By : Dipti
Date : 18.11.2024

Checked by : yogita
Date : 18.11.2024