


INITIAL DISINTEGRATION TEST REPORT

Product Name	HardGenlatin Tablet		
Batch No.	BT-003	Department	NA
BFG Code No.	BFG CODE NO-012	Date	06.04.2024
No. of Samples	5	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	EQ Code No-1	Disintegration Tester Code No.	ED 3PO
Media	NA	Side	NA
Temperature Range (°C)	35.0 - 38.0		

Test Results

Min. Temperature (°C)	28.6	Max. Temperature (°C)	28.6
Max. Disintegration Time	00:01:39	Set Temperature	38.0
Mode	NA	Jar Detail	A
Basket Type	10 Mesh		



Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	15:05:31	15:06:42	NA	00:02:00	00:01:11

Result: **Not Complies**
test

Performed By : Tantra
Date : 06.04.2024

Checked by : Mangesh
Date : 06.04.2024