


INITIAL DISINTEGRATION TEST REPORT

Product Name	captab capsules		
Batch No.	BT-CAPS-041	Department	QA
BFG Code No.	bfg-002	Date	19.08.2025
No. of Samples	5	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	Equipment-123	Disintegration Tester Code No.	DT-EDI-2SA
Media	CO2	Side	NA
Temperature Range (°C)	32.0 - 38.0		

Test Results

Min. Temperature (°C)	35.2	Max. Temperature (°C)	39.7
Max. Disintegration Time (HH:MM:SS)	00:14:59	Set Temperature (°C)	37.0
Mode	Dual	Jar Detail	B
Basket Type	NA		



Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	15:12:45	15:27:44	NA	00:01:28	00:14:59

Result: **Not Complies**

Remark: TEST 1234567890POIUYTREWQ1234567890POIUYTREWQWERTYUIUYTREWQQWERTYUIUYTEWERY

Performed By : kuldeep

Checked by : Mangesh

Date : 19.08.2025 & 15:12:45

Date : 19.08.2025 & 16:52:37