


REGULAR DISINTEGRATION TEST REPORT

Product Name	KROCIN TABLET1		
Batch No.	BT-COAT-0111DT	Department	QC
BFG Code No.	KROCIN TABLET1	Date	30.07.2025
No. of Samples	5	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	None	Disintegration Tester Code No.	DT-ED 3PO
Media	CO2	Lot No.	45g
Temperature Range (°C)	35.0 - 39.5		

Test Results

Min. Temperature (°C)	39.6	Max. Temperature (°C)	39.6
Max. Disintegration Time (HH:MM:SS)	00:12:05	Set Temperature (°C)	38.0
Mode	NA	Jar Detail	B
Basket Type	10 Mesh		



—●— Actual Value —●— Standard

Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	15:08:32	15:20:37	NA	00:03:05	00:12:05

Result: **Not Complies**

Remark: TEST

Performed By : kuldeep

Date : 30.07.2025 & 15:08:32

Checked by : Dipti

Date : 30.07.2025 & 11:24:46