


REGULAR DISINTEGRATION TEST REPORT

Product Name	DOLO 500		
Batch No.	BT-DT	Department	QC
BFG Code No.	456	Date	06.01.2026
No. of Samples	5	Std. Disintegration Time (NMT)	00:15:02
Equipment Code No.	None	Disintegration Tester Code No.	EDI-2SA
Media	H2O	Side	RHS
Temperature Range (°C)	35.5 - 39.5	Stage	Compression

Test Results

Min. Temperature (°C)	37.8	Max. Temperature (°C)	38.6
Max. Disintegration Time (HH:MM:SS)	00:14:59	Set Temperature (°C)	37.2
Mode	Dual	Jar Detail	B
Basket Type	10 Mesh		



Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	15:05:26	15:20:25	RHS	00:00:00	00:14:59

Result: **Complies**

Performed By : yogita
Date & Time : 06.01.2026 & 15:05:26