


REGULAR DISINTEGRATION TEST REPORT

Product Name	HardGenlatin Capsule		
Batch No.	125	Department	QC
BFG Code No.	BFG CODE NO-123	Date	12.04.2024
No. of Samples	0	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	Equipment-123	Disintegration Tester Code No.	ED 3PO
Media	NA	Side	NA
Temperature Range (°C)	35.0 - 38.5		

Test Results

Min. Temperature (°C)	35.6	Max. Temperature (°C)	35.6
Max. Disintegration Time	00:15:00	Set Temperature	38.0
Mode	NA	Jar Detail	A
Basket Type	10 Mesh		



Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	15:05:31	15:20:31	NA	00:15:05	00:15:00

Result: **Complies**

Performed By : Tantra
Date : 12.04.2024