


INITIAL DISINTEGRATION TEST REPORT

Product Name	PARACETAMOL_15 MG		
Batch No.	BT-SHIFT-1	Department	QC
BFG Code No.	PARA_123	Date	08.05.2024
No. of Samples	5	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	EQ001	Disintegration Tester Code No.	ED-3PO
Media	NA	Side	NA
Temperature Range (°C)	35.0 - 39.0		

Test Results

Min. Temperature (°C)	37.0	Max. Temperature (°C)	37.0
Max. Disintegration Time	00:01:11	Set Temperature	38.0
Mode	NA	Jar Detail	B
Basket Type	10 Mesh		



Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	15:08:32	15:09:43	NA	00:00:00	00:01:11

Result: **Complies**

Performed By : Abhishek
Date : 08.05.2024