


REGULAR DISINTEGRATION TEST REPORT

Product Name	KROCIN TABLET1		
Batch No.	BT-DT Summary C	Department	QC
BFG Code No.	KROCIN TABLET1	Date	12.11.2024
No. of Samples	10	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	EQ 1	Disintegration Tester Code No.	ED 3PO
Media	WATER	Lot No.	r
Temperature Range (°C)	35.0 - 39.5		

Test Results

Min. Temperature (°C)	36.7	Max. Temperature (°C)	36.7
Max. Disintegration Time	00:14:39	Set Temperature	38.0
Mode	NA	Jar Detail	C
Basket Type	10 Mesh		



Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	15:12:01	15:26:40	NA	00:00:16	00:14:39

Result: **Complies**

Performed By : Dipti

Date : 12.11.2024