


REGULAR HARDNESS & DIMENSION REPORT

Product Name	crocin		
Batch No.	hrdn-88	Department	QC
BFG Code No.	only hrdnes	Date	30.12.2025
No. of Samples	5	Start Time	10:27:16
Equipment Code No.	EQP-0920	Hardness Tester Code No.	HRD_25
Side	NA		
Stage	Compression		

Standard Limits					
Hardness (Kp)		Thickness (mm)		Length(mm)	
Lower =	50	Lower =	0	Lower =	0
Standard =	100	Standard =	0	Standard =	0
Upper =	150	Upper =	0	Upper =	0



REGULAR HARDNESS & DIMENSION REPORT

Test Results			
Sample No.	Hardness (Kp)	Thickness (mm)	Length(mm)
1	50	NA	NA
2	50	NA	NA
3	50	NA	NA
4	50	NA	NA
5	50	NA	NA
Statistics			
Average	= 50.0 Kp	Average	= NA
Minimum	= 50 Kp	Minimum	= NA
Maximum	= 50 Kp	Maximum	= NA
Std. Deviation	= 0.0000	Std. Deviation	= NA
Minimum %	= 50.00	Minimum %	= NA
Maximum %	= 50.00	Maximum %	= NA
Result:	Complies	Result:	NA

Performed by : yogita

Date & Time : 30.12.2025 & 10:27:16