

BULK / TAPPED DENSITY TEST REPORT

Product Name	PARACETAMOL_50 MG		
Batch No.	BT-IPQCTD-1036	Department	QC
BFG Code No.	PARA_TAB001	Date	04.09.2025
Method	USP II	Start Time	14:36:40
Bulk Std. Lower Limit (g/ml)	0.10	Bulk Std. Upper Limit (g/ml)	0.70
Tapped Std. Lower Limit (g/ml)	0.50	Tapped Std. Upper Limit (g/ml)	0.90
Tapped Density Apparatus Code No.	TD002		

1 .BULK DENSITY :

Quantity of the test sample taken(A) : 55.0 g

The volume occupied(Vo) : 80.0 ml

$$\text{Bulk Density} = \frac{\text{Quantity of the test sample}}{(\text{Vo})} = \frac{55.0}{80.0} = 0.688 \text{ g/ml}$$

2 .TAPPED DENSITY :

Quantity of the test sample taken(A) : 55.0 g

Number of tap count set : 10 + 500 + 1250 + 1250

The tapped volume (Va) : 70.0 ml (Volume after 10 tapping)

(Vb) : 68.0 ml (Volume after 500 Additional tapping)

(Vc) : 67.5 ml (Volume after 1250 Additional tapping)

(Vd) : 66.0 ml (Volume after 1250 Additional tapping)

* Note: If difference between Vb and Vc exceeds 2 ml ,repeat 1250 Tapping.

Additional 1250 taps (V1250b) : 65.0 ml

$$\text{Tapped Density} = \frac{\text{Quantity of the test sample taken}}{(\text{Vc}) \text{ or } (\text{Vd})} = \frac{55.0}{65.0} = 0.846 \text{ g/ml}$$

Result: **Complies**
Remark: TEST TTTTTTTTTTTTTTTTEEEEEEEEEEESSSSSSSSSSSSSSTTTTTTTTTTTTTTTTTTT

Performed by : kuldeep

Date : 04.09.2025 & 14:36:40

Checked by : Mangesh

Date : 04.09.2025 & 14:37:03