



BATCH REPORT OF FRIABILITY TEST
Area : Compression

Product Name	KROCIN TABLET1		
Batch No.	BT-Summary Doub	Department	QC
BFG Code No.	KROCIN TABLET1	Std. Limit NMT(%W/W)	1.00

Batch Summary		
Minimum Friability(%w/w)	Maximum Friability(%w/w)	Average Friability
NaN	NaN	NA



Detailed Report							
Test No.	Date	Time	Friabilator Code No.	Side	Friability (%w/w)	Test Result	Performed By
1	14.11.2024	12:47:15	FRI002	LHS	1.00	Complies	Dipti
2	14.11.2024	12:47:22	FRI002	RHS	0.76	Complies	Dipti
3	14.11.2024	12:53:31	FRI002	LHS	0.37	Complies	Dipti
4	14.11.2024	12:53:46	FRI002	RHS	1.76	Not Complies	Dipti

BATCH REPORT OF FRIABILITY TEST
Area : Compression

Test No.	Date	Time	Friabilator Code No.	Side	Friability (%w/w)	Test Result	Performed By
5	14.11.2024	13:44:38	FRI002	RHS	NaN	Not Complies	Dipti
6	14.11.2024	13:44:47	FRI002	RHS	1.00	Complies	Dipti
7	14.11.2024	13:47:24	FRI002	RHS	NaN	Complies	Dipti
8	14.11.2024	13:47:24	FRI002	RHS	0.02	Complies	Dipti
9	14.11.2024	15:16:48	FRI002	RHS	NaN	Not Complies	Dipti
10	14.11.2024	15:17:06	FRI002	RHS	1.96	Not Complies	Dipti