


DISINTEGRATION TEST REPORT

REPRINT

Product Name	FENOFIBRATE TABLETS USP 145 MG		
Batch No.	IDDT-Double	Department	Compression
BFG Code No.	31004737	Date	15.04.2022
No. of Samples	NA	Std. Disintegration Time (NMT)	00:15:00
Equipment Code No.	T/270	Disintegration Tester Code No.	hhh
Media	0.1 M HCL (GB)	Side	LHS
Temperature Range (°C)	36.0 – 38.0		

Test Results

Min. Temperature Jar A/B(°C)	36.7	Max. Temperature Jar A/B (°C)	37.7
Max. Disintegration Time LHS	00:04:07		



Sample No.	Start Time	End Time	Side	Halt Duration	Disintegration Time
1	16:29:56	16:30:56	LHS	00:00:08	00:01:00
2	16:29:56	16:34:03	LHS	00:00:08	00:04:07
3	16:29:56	16:30:08	LHS	00:00:08	00:00:12
4	16:29:56	16:31:11	LHS	00:00:08	00:01:15
5	16:29:56	16:32:04	LHS	00:00:08	00:02:08
6	16:29:56	16:31:21	LHS	00:00:08	00:01:25

Result: **Complies**

Performed by : Heena Mewada

Date : 15.04.2022