



Product Name	CTNGlimy2mgTab5x2x15'sDomR		
Batch No.	Batch 5896	Repetition No.	0
Product Code	150098128	Quantity	10
SFOID	NA	Date	03.09.2024
Equipment ID	HR Equipment	Time	10:14:19
Side	LHS	Hardness Code	MB-Hardness-9875
Serial No.	18		

Standard Limits					
Hardness (N)		Thickness (mm)		Length(mm)	
Lower Limit =	0.63	Lower Limit =	0.14	Lower Limit =	0
Upper Limit =	8.63	Upper Limit =	0.79	Upper Limit =	0



Done By : hasanraza(55)

Printed By : Nirupam Giri(3)

Date & Time : 03.09.2024 10:14:19

Date & Time : 03.09.2024 10:14:51

This document is Confidential, electronically generated by Nirupam Giri 03.09.2024 10:14:51

Product Name	CTNGLimy2mgTab5x2x15'sDomR		
Batch No.	Batch 5896	Repetition No.	0
Product Code	150098128	Quantity	10
SFOID	NA	Date	03.09.2024
Equipment ID	HR Equipment	Time	10:14:19
Side	LHS	Hardness Code	MB-Hardness-9875

Test Results			
Sample No.	Hardness (N)	Thickness (mm)	Length(mm)
1	7	0.60	0
2	7	0.60	0
3	7	0.60	0
4	7	0.60	0
5	7	0.60	0
6	7	0.60	0
7	7	0.60	0
8	7	0.60	0
9	7	0.60	0
10	7	0.60	0
Statistics			
Average = 7	Average = 0.60	Average = NA	
Minimum = 7	Minimum = 0.60	Minimum = NA	
Maximum = 7	Maximum = 0.60	Maximum = NA	
Std. Deviation = 0.0	Std. Deviation = 0.00	Std. Deviation = NA	
Remark: Within Limit	Remark: Within Limit	Remark: NA	

Done By : hasanraza(55)

Printed By : Nirupam Giri(3)

Date & Time : 03.09.2024 10:14:19

Date & Time : 03.09.2024 10:14:51

This document is Confidential, electronically generated by Nirupam Giri 03.09.2024 10:14:51