

LOSS ON DRYING (LOD) TEST RECORD
(GRANULATION)

Product / Method Name	Esom EC 40 mg + Dom SR 30m		
Date	17.01.2025	Product Code	200101341
Batch No.	NTT6	MPR No.	cpg-t19
SFOID	1213	Make	Mettler Toledo
Model	HX204	Instrument ID	PR703
Serial No.	C004959123	Repetition No.	22
End Time	15:43	Switch Off Mode	5:00 min
Total Time	5:00 min	Drying Program	Standard
Loss On Drying Range	NMT 3.5 % w/w (Target of 2.5% w/w)	Dry Temperature (°C)	105

Test Results	
Initial Weight (g)	2.635
Dry Weight (g)	2.546
End Result (%MC)	3.510

Result: **Fail**

Performed by : Jaisukh BK

Date : 17.01.2025