

VACUUM LEAK TEST RECORD
(PACKING)

Product Name	FLUTICASONE		
Batch No.	Batch-2222	Repetition No.	NA
Product Code	NA	Date	21.09.2024
SFOID	SFO-2222	Start Time	15:29:15
Instrument Id.	DP310824	Test ID	18092417190322
Batch Status	Initial Setup	Model No.	ELT-201WP
Serial No.	18	MPR No.	MPN_2

Test Details

Dosage Units	4	Dosage Form	Liquid
Package Type	Bottle	EORT Vacuum Level	NA
Vacuum Build Time	NA	Vacuum Release Time	NA
Result	Visual Inspection	Test Status	Aborted

Method Details

Method Name	FLUTICASONE	Vacuum Level	400 mmHg
No. of Packages	4	Desicator Size	Small
Vacuum Retention Time	03:00	Vacuum Penetration Time	01:00

Result: _____

Performed by : Ashwin Reddy

Date : 21.09.2024