

Print Date : 07.01.2025

Print Time : 17:16:27

Print By : Initial User

Product Name : Esom EC 40 mg + Dom SR 30m

DR.REDDY'S LABORATORIES LTD

FTO Unit-3

EQ001 - Batch Report DT (mm:ss)

Batch Date : 2025-07-01

Batch No : R33

|                          |             |
|--------------------------|-------------|
| Product Code             | 200101341   |
| Compression Equipment ID | EQ001       |
| Type Of Rotary           | Double      |
| Area                     | Compression |
| Disintegration tester ID | PRE1054     |
| No Of stations           | 12          |

|                  |                     |
|------------------|---------------------|
| Batch Start Time | 07.01.2025 16:22:29 |
| Batch Stop Time  | 2025-01-07 04:23:32 |
| UserName         | Initial User        |

| S. No. | Start Time          | Stop Time           | Instrument ID | LHS Container No | RHS Container No | NMT( mm:ss ) | Min   | Max   | Avg   | Result | UserName   |
|--------|---------------------|---------------------|---------------|------------------|------------------|--------------|-------|-------|-------|--------|------------|
| 1      | 07.01.2025 16:22:29 | 07.01.2025 16:22:29 | PRE1054       | 342              | NA               | NA           | 12:08 | 12:08 | 12:08 | PASS   | Jaisukh BK |
| 2      | 07.01.2025 16:22:29 | 07.01.2025 16:22:29 | PRE1054       | 342              | NA               | NA           | 12:08 | 12:08 | 12:08 | PASS   | Jaisukh BK |

|            |        |
|------------|--------|
| Avg Of Avg | 12:080 |
|------------|--------|

Checked By :

\*NA-Not applicable

Print Date : 07.01.2025  
Print Time : 17:16:27  
Print By : Initial User  
Product Name : Esom EC 40 mg + Dom SR 30m

DR.REDDY'S LABORATORIES LTD  
FTO Unit-3  
EQ001 - Batch Report DT (mm:ss)

Batch Date : 2025-07-01  
Batch No : R33

