

pH Calibration Report

Reprint

Thermo Scientific (c) 2011

Versa Star Pro

Meter S/N	V122318
SW Rev	11.50
Channel	1
Module S/n	VA32174
Method	100
Calibration Time	17-09-24,15:07

-- pH Calibration Report --

Point	1
pH	7.00
mV	0.0
Temperature	25.0 C (MAN)
Calibration Type	Auto
Average Slope	100.0 %
Cal #	103

Meter ID : QC-PHMTZ-009

Probe ID : CX1-16015

Tolerance : 90.00 % - 105.00 %

Result : Pass

Done By : Trident

Reviewed By : Trident

Comment : As part of PQ

Date & Time : 17-09-2024 & 15:33:46

Printed By : Thermofisher

Date & Time : 16-01-2025 & 11:27:00