



pH Calibration Report

Reprint

Thermo Scientific (c) 2011

Versa Star Pro

Meter S/N V122320
SW Rev 11.50
Channel 1
Module S/n VA36002
Method 100
Calibration Time 13-09-24,09:35

-- pH Calibration Report --

Point 1
pH 7.00
mV 0.0
Temperature 25.0 C (MAN)
Calibration Type Auto
Average Slope 100.3 %
Cal # 103

Meter ID : QC-PHMTZ-007

Probe ID : CX1-15733

Tolerance : 90.00 % - 105.00 %

Result : Pass

Done By : Trident

Reviewed By : Trident

Comment : As part of OQ

Date & Time : 13-09-2024 & 10:05:07

Printed By : Thermofisher

Date & Time : 02-06-2025 & 09:22:03